

CERTIFICATE OF ANALYSIS

Order Type: Medical Cannabis

Customer ID: 32

Harvest/Extract Lot: SL + BS10000
040119

Order ID: OR2019-939

Customer Name: CBD Plus USA

Harvest/Extract Batch: GP 2121
040119

Cultivar (Strain): 1000mg BS
Strawberry Lemonade

Lab ID: SA2019-3244

Sample Matrix: Oil/Tincture

Sample Date: 04/03/2019

Date Received: 04/03/2019

Date Completed: 04/05/2019

Remarks:

CANNABINOID (POTENCY) PROFILE

Analysis Date/Time: 04/04/2019 0706

Method: HPLC/DAD (Internal Method-001)

Moisture Content (%): -

Analyst: OL

Instrument: Agilent 1100

Water Activity (aw): -

<u>Cannabinoid</u>	<u>Result (%)</u>	<u>Result (mg/mL)</u>	<u>Reporting Limit (mg/mL)</u>	<u>Per Unit (mg)</u>
CBD	3.33	33.3	0.0025	998
CBDa	-	-	0.0025	-
CBDv	0.0101	0.101	0.0025	3
Δ9-THC	-	-	0.0025	-
Δ8-THC	-	-	0.0025	-
THCa	-	-	0.0025	-
THCv	-	-	0.0025	-
CBC	0.0692	0.692	0.0025	21
CBG	0.0154	0.154	0.0025	5
CBGa	-	-	0.0025	-
CBN	0.0265	0.265	0.0025	8




TOTAL	3.45	34.5		1035
TOTAL THC	-	-		-
TOTAL CBD	3.33	33.3		998

UNIT VOLUME (mL): 30

"-" Not detected above RL.

Cannabinoid Distribution
(% of Total Cannabinoids)



Deviations from standard operating procedure: None

Recoveries for all analyte standards: 90-110%
Replicate Uncertainties: <5% RSD, <20% RPD
Sample/Reagent Blanks: <RL for all analytes

Values for plant matter are adjusted for moisture content.

Total THC = (THCa x 0.877) + Δ9-THC
Total CBD = (CBDa x 0.877) + CBD

Percentage results are reported by mass.
mg/g results are reported as mass component per mass material.

Abbreviations: UV - Ultraviolet, HPLC - High Pressure Liquid Chromatography, RL - Reporting Limit, RPD - Relative Percent Difference, RSD - Relative Standard

This information is provided as a service and makes no claims of efficacy and/or safety of this product. Results are applicable only for the sample(s) analyzed and for the specific analysis conducted. This report is for informational purposes only and should not be used to diagnose, treat, or prevent any medical-related symptoms.

REPORT OF LABORATORY ANALYSIS

herein have not been approved and/or endorsed by the manufacturer, except in full, without the written consent of Felling Analytical Services and Technology (F.A.S.T.), LLC

Kyle W. Felling
Kyle W. Felling, Ph.D.
Laboratory Director